# **ARJOHUNTLEIGH**

**GETINGE GROUP** 

#### **EASYTRACK FS REP**

MODEL №: 9720090 Instruction for Use



( (

...with people in mind



### **Table of Contents**

Foreword	1
Thank you for choosing ArjoHuntleigh products	1
Please Read this Manual Thoroughly!	1 1
Manufacturer Information	
Definitions Used in this Manual	
Safety Instructions	2
Intended Use	
Operational Life	
How to Use this ManualSafe Working Load	
Important Safety Instructions	
Environmental Advice	
Homecare Environment Considerations	
Symbols Used	
General Symbols	
Product Description / Labels	5
EASYTRACK FS REP System Requirements	6
Floor Requirements:	
Compatibility	6
Room Assessment	7
Assembling the EASYTRACK FS REP	8
Step 1. Positionning the Posts	
Step 2. Installing the Track	
Step 3. Adjusting Track WidthStep 4. Preparing the Ceiling Lift	
Step 5. Replacing the Lift/Trolley onto the Track	. 10
Step 6. Inserting the Lift onto the Track	. 10
Step 7. Installing the Charging Station	
Extending the Posts	
Levelling the Track	
Retracting the Posts	
Securing the Posts	
Levelling the Posts	
Disassembling the EASYTRACK FS REP	
Accessories	.12
Care and Maintenance	.13
Preventive Maintenance Schedule	
Annual Inspection by an Authorized Technician	
· · · · · · · · · · · · · · · · · · ·	
Troubleshooting	
Technical Specifications	
Dimensions	. 15

### Thank you for choosing ArjoHuntleigh products.

Your EASYTRACK FS REP™ is part of a series of quality products designed specially for home care, nursing homes and other health care uses.

We are dedicated to serving your needs and providing the best products available along with training that will bring your staff maximum benefit from every ArjoHuntleigh product.

#### Please Read this Manual Thoroughly!

The information in this manual is crucial to the proper use and maintenance of the *EASYTRACK FS REP*. It will help protect your product as well as ensure that it performs to your satisfaction.

Lifting and transferring a person always presents a potential risk. This manual contains safety related information that must be read and understood to help prevent injuries.

**WARNING**: ArjoHuntleigh strongly advises and warns that to avoid injuries that can be attributed to the use of inadequate parts, only parts designated by ArjoHuntleigh should be used on product and other appliances supplied by ArjoHuntleigh.

**WARNING**: Unauthorized modifications on any ArjoHuntleigh product may affect its safety. ArjoHuntleigh will not be held responsible for any accidents, incidents or lack of performance that occur as a result of any unauthorized modification to its products.

Do not attempt to service or repair the *EASYTRACK FS REP*. If any item is defective, contact your local ArjoHuntleigh vendor.

#### **Service and Support**

A service routine has to be done on your *EASYTRACK FS REP* by ArjoHuntleigh trained service personnel. This will ensure it remains safe and functional.

Please contact your local ArjoHuntleigh vendor for any of the following:

- If you require further information.
- Want to report an unexpected event, change in the performance or a malfunction.
- Need any help in setting up, using or maintaining your EASYTRACK FS REP.
- · Need replacement parts.

Your ArjoHuntleigh vendor can offer support and service programs to maximize the longterm safety, reliability and value of the product.

Additional copies of this manual can be purchased from your local ArjoHuntleigh agent. When ordering, include the Instructions for Use part number (see front page).

#### **Manufacturer Information**

This product was manufactured by:

ArjoHuntleigh AB Verkstadsvägen 5 241 38 Eslöv SWEDEN

☎: +46 (0) 10-335 45 00 畳: +46 (0) 413-138 76 ❤: www.ArjoHuntleigh.com

#### **Definitions Used in this Manual**

#### WARNING:

Means: Failure to understand and follow this instruction may result in injury to yourself and others.

#### **CAUTION:**

Means: Failure to follow this instruction may cause damage to the product(s).

#### NOTE:

Means: Important information regarding correct use of the product.

### **Safety Instructions**

#### **Intended Use**

The EASYTRACK FSREP is a portable track system to be used for demonstration, with ArjoHuntleigh ceiling lift, as an assistance device for transferring a person between the posts. It can be used in a homecare setting, at nursing homes and other assisted living centres.

Patient transferring is performed under the supervision of a trained caregiver as per the instructions outlined in this manual in conjunction with the *Instruction for Use* of the lift attached to the *EASYTRACK FS REP*. All other uses must be avoided.

The EASYTRACK FS REP must only be used for the purposes stated above.

**WARNING**: The *EASYTRACK FS REP* is for transferring patients only. Do not use the *EASYTRACK FS REP* for any other purpose.

Using this *EASYTRACK FS REP* outside the scope of its intended use could lead to injuries.

#### **Operational Life**

This product is designed and tested for a useful life of seven (7) years or 10000 transfers — whichever comes first—at maximum safe working load (SWL). It is subject to preventive maintenance as specified in the "Care and Maintenance" section.

Time equivalence between the number of transfers versus the number of years is made clear in the following table.

Transfers per Day	Years (10 000 transfers)
4	7
6	4.5
8	3.5

**WARNING**: The manufacturer cannot ensure full safety for a product of which the life span has been exceeded. Wear may cause the breakage of a part and lead to a patient fall.

The life span is the safe operating time period before a complete overhaul is required.

Following are factors that have an impact on the *EASYTRACK FS REP* life span:

- aging of the unit;
- transfers per day;
- weight of the patient;
- maintenance frequency.

#### **How to Use this Manual**

To ensure the safe operation of the *EASYTRACK FS REP* read the entire manual carefully before operating this product.

If anything is not completely understood, please contact your local ArjoHuntleigh agent for more details. Failure to comply with warnings in this manual may result in injuries.

Keep this manual handy and refer to it as required. Make sure that all operators are regularly trained in the use of the product as per the information found therein.

The EASYTRACK FS REP must be used as per these safety instructions.

Anyone using the EASYTRACK FS REP must also have read and understood the instructions in this manual.

If there is anything you are not sure about, contact your local ArjoHuntleigh agent.

#### Safe Working Load

The EASYTRACK FS REP has been designed to be used with ArjoHuntleigh lifts with a safe working load of 200 kg (440 lb) or less.

**WARNING**: The *EASYTRACK FS REP* is intended to be used for residents whose weight is within a specified safe working load (SWL). Do not attempt to lift more than the lowest weight limit indicated on the following:

- the "maximum load" label on the lift
- on the spreader bar
- on the sling.

Overloading any of these elements could lead to a patient fall.

#### **Important Safety Instructions**

Before installing or using the *EASYTRACK FS REP* always ensure to comply with following safety instructions.

- Carefully inspect product upon arrival. If any pieces are missing or appear damaged – do not install. Contact your local ArjoHuntleigh agent for further instructions.
- All labels are read and those that cannot be easily read are replaced.
- The daily maintenance is carried out before using the unit.
- Do not drop product; may cause breakage.
- Do not submerge in water.
- Do not attempt to over-extend the track past the locking point.
- Do not bump the posts This might cause unit to become unstable. If posts are bumped, inspect the EASYTRACK FS REP system to ensure the posts are straight.
- Never use the EASYTRACK FS REP system as a swing.
- Do not allow the patient to swing in any direction while being lifted.
- Do not lean against posts.

**WARNING:** Never use, store or carry the *EASYTRACK FS REP* when a red marker is visible. Serious injuries could occur. Refer to "Fig.1" on page 5 for location of red markers.

- Protect the EASYTRACK FS REP and its accessories during transport.
- Prior each use, perform the "Before every use" inspection points in the "Preventive Maintenance Schedule" table on page 13.
- Always make sure that the posts are fully retracted when transporting or storing.

#### **Environmental Advice**

When disposing of any items associated with the product, contact local authorities for information.

### Homecare Environment Considerations

**WARNING:** The *EASYTRACK FS REP* is not intended to be operated by children. Serious injuries could occur.

**NOTE:** Rigorous cleaning actions should be done when the *EASYTRACK FS REP* is exposed to an animal. Pet hair trapped inside the device can reduce the product performance.

## Symbols Used

### **General Symbols**

	Manufacturer's name and address. May also point out manufacturing date.
C€	Product complies with the medical device directive 93/42/EEC.
REF	Points out the manufacturer's catalogue number.
$\triangle$	Indicates a potential risk, refer to <i>Instructions for Use</i> (IFU) for details.
SN	Points out the manufacturer's serial number.
SWL	SWL is the maximum load the device is rated for safe usage.
(S)	Refer to the Instructions for Use.
0	Incorrect
0	Correct

### **Product Description / Labels**




Fig.1

Item	Part Description	Quantity
1	Base	2
2	Extendable post	2
3	Extendable track with trolley	1
4	Ceiling lift guide	2
5	Extension cable	1
6	Clip-on charging station	1
Itom	Label Description	Quantity

Item	Label Description	Quantity
Α	Product Identification, Manufacturer Data	1
В	Warning	4
С	Product Name	4
D	Maximum Load Capacity (SWL)	2
Е	Track Identification, Manufacturer Data, SWL	1
F	FS Logo	4
G	Anti-slip Sticker	2
Red N	Markers - Do not use the EASYTRACK FS REP when any of the following marker	is visible
Н	Visible when inner track is extended beyond maximum usage lenght	4
J	Visible when collar clip is unlocked	8
K	Visible when treak heak is unlatehed	4
L	Visible when track hook is unlatched	4
М	Visible when treek heak letch handle is unlesked	4
N	Visible when track hook latch handle is unlocked	4

### EASYTRACK FS REP System Requirements

#### Floor Requirements:

**CAUTION:** If you are unsure of the strength or stability of the floor – do not install the *EASYTRACK FS REP* System – consult your local ArjoHuntleigh agent.

Flooring must comply to local building code standards.

WARNING: Termites or water damages may affect the structure of the floor.

- Floor must be clear of any dust, grease, water or any foreign substance.
- Floor must have a slope of less than 0.5°. See drawing below to determine the floor slope. Use a level or contact a contractor if unsure of floor slope.




Fig.2

#### **Compatibility**

The EASYTRACK FS REP can only be used with the following ceiling lifts.

Lift Name	Product №	Category	Remarks
Voyager Duo	913XXXX	Fixed	
V5 Duo	923XXXX	Fixed	
Voyager	980X00.XX	Portable	Includes Voyager Lite
Maxi Sky 600	LDXXXXX-XX	Fixed	
Maxi Sky 440	LEXXXXX	Portable	
Maxi Sky 2	MSXXX-XX-XX	Fixed	

Note: Fixed ceiling lift must feature holes on the top housing to allow installing the track guides.

### **Room Assessment**

Before installing the *EASYTRACK FS REP* prepare the installation as follows:

- 1) Survey the room where vou plan to install the EASYTRACK FS REP system.
  - Make sure the room meets all the specifications outlined in "EASYTRACK FS REP System Requirements"
- 2) Make sure the pick-up point will be inside the transfer zone:
  - Over a bed : approximately 915 - 1020 mm (36 - 40") from the head of the bed.
  - Over a tub: approximately 400 - 460 mm (16 - 18") from the end of the
  - Allow 1000 mm (39") of the open area for transfer working space.




Fig. 3

3) Determine the location for the two posts considering that transfers should only take place between the posts of the *EASYTRACK FS REP* system.




Fig.4

### Assembling the EASYTRACK FS REP

**NOTE:** The assembly can be performed by one person. Perhaps, we recommend to be assisted for increased safety.

**WARNING**: Never leave a post unattended when the track is not assembled. The post could fall down resulting in injuries.

#### **Step 1. Positionning the Posts**

- Remove all parts from the boxes. Before assembling make sure that all parts are included and in good condition.
- 2) Layout the bases on a level and clean floor at the desired distance.
- 3) Insert the posts in the bases. The design of the base will ensure the post lines up in the correct position.



Fig.5

**CAUTION:** Make sure both posts are correctly inserted into each base.




Fig.6

#### **Step 2. Installing the Track**

 Prior to track installation, unlock the hook handle by pressing on the safety latch and pulling outward (located on the blue colour component on the side of post head).



Fig.7

2) Gently pull out the handle until the red stripe of the hook mechanism is visible.



Fig.8

Insert the track in place on top of both posts.



Fig.9

 Attach the track to the post by engaging the hooks, on both sides of the post, over the track pins.



Fig. 10



Fig. 11

You should feel a slight tension when closing the hook handle.

Repeat procedure until all 4 hooks are secured.

Following figure shows a view of the track when properly installed and locked in place.



Fig. 12

**CAUTION:** Make sure the main hooks are well engaged.

### Step 3. Adjusting Track Width

 To adjust the track width, pull the inner track away from the main outer track until the desired width is achieved.




Fig. 13

2) Trolley position should be as shown in "Fig. 13". If not, make sure that inner track is extended. Trolley will become free of its transportation slot. If it does not, shake the trolley lightly left and right to release it. Trolley should move freely at this point.

**NOTE:** This step can also be done after adjusting the height.

#### Step 4. Preparing the Ceiling Lift

This section only applies if you expect to insert a fixed ceiling lift onto the track.

The ceiling lift needs the following modifications prior being inserted onto the track. These will allow correct alignment with the charging station (all models). It will also prevent the housing from interfering with the track on Maxi Sky 600 model.

1) Install the 2 track guides on top of the lift's housing.




Fig. 14

**Note:** Older Maxi Sky 600 housings may not feature holes allowing to install the guides. You may need to change the housing for a newer version.

2) For MaxiSky 600 only: Remove the idler gear protectors.





Fig. 15

### Step 5. Replacing the Lift/Trolley onto the Track

1) Loosen the special knob on top of the track until the mark on the thread is visible.



2) Press on the extension stopper to allow the inner track to slide out of the outer track.




Fig. 17

 Slide the track out until the red stripe is entirely visible. This will open an access under the track allowing to change the lift/ trolley onto the track.




Fig. 18

### Step 6. Inserting the Lift onto the Track

- Insert the lift through the track openning located below the red stripe.
- 2) Slide the lift toward opposite end of the track.
- Retract the inner track. You should hear a click indicating that the track stopper is properly re-engaged.
- 4) Adjust the length of the track (extendable from either end).
- 5) Gently tighten down the special knob.

### Step 7. Installing the Charging Station

This section only applies for fixed ceiling lifts.




Fig. 19

- 1) Locate the side of the track where the lift contact blades are placed.
- 2) Clip the charging station on that side of the track.
- Route the cabling onto the upper channel of the track. Use the extension to allow the cable to reach an AC outlet and connect the charger.

#### **Height Adjustment**

#### Extending the Posts

 Pull on the lower portions of each collar locking handle located on both sides of the collar to release the locking mechanisms and allow the inside section of the post to extend upwards.



Fig. 20

When the clamping collars are in the open position red marks are visible.

 Position yourself under the track and gently push upwards until the desired height is achieved.



Fig.21

**CAUTION:** NEVER extend the post to allow the track to be in contact with the ceiling. Always leave a clearance of at least 30 mm (1 inch) above the track. This clearance is necessary to allow retracting the post for disassembling.

#### Levelling the Track

4) Ensure the track is level. To level the track, extend both posts to same length. Make sure the ruler in front of the post is exiting the post by same lenght for both posts.



Fig.22

#### Retracting the Posts

Proceed as follows to retract the post for adjusting the length or for disassembling.

- 5) Unlock the collar of the post, if it is not already unlocked.
- 6) Refering to "Fig.23", grab the upper portion of the post with both hands (1) and raise it about 5 mm to remove the pressure on the collar.
  - Take care not to allow the track to make contact with the ceiling.
- 7) With your two thumbs, slide only the ruler downward (2) then lower the post to desired height (3).




Fig.23

8) When the ruler is released, it reverts back to its original position preventing the post from being contracted.

#### Securing the Posts

 Lock the clamping collars on both posts to prevent the posts from extending by themselves. When the collars are locked, no red mark should be visible.

#### Levelling the Posts

10) Adjust the levellers on each base to stabilize the bases. When adjusting the levellers, make sure the central gliders are in contact with the floor.




Fig.24v


Your EASYTRACK FS REP is now ready to be used with your ArjoHuntleigh portable lift (not included).

### Disassembling the EASYTRACK FS REP

To disassemble the *EASYTRACK FS REP* reverse the assembly sequence.

- 1) Remove the lift from the trolley.
- 2) Unlock the clamping collars on both posts.
- 3) Retract both posts (see "Retracting the Posts" on page 11 and relock the clamping collars.
- 4) Unlatch the hook handles to release the hooks from the track pins.
- 5) Remove the track from the posts and retract it.
- 6) Remove both posts from their bases.
- 7) Place all components in an appropriate container to protect them from dust or damages.





#### **Preventive Maintenance Schedule**

The EASYTRACK FS REP is subject to wear and tear, and the following actions must be performed when specified to ensure that the product remains within its original manufacturing specification.

WARNING: The points on this checklist are the minimum the manufacturer recommends. In some cases more frequent inspections should be carried out. Continuing to use the EASYTRACK FS REP without conducting regular inspections will seriously compromise the user and resident / patient's safety. Preventive maintenance specified in this manual can prevent accidents.

WARNING: Safety related maintenance and authorized service must be carried out by qualified personnel, fully trained in servicing procedures by ArjoHuntleigh and equipped with correct tools and proper documentation, including Parts List and Service Manual. Failure to meet these requirements could result in personal injuries and/or unsafe equipment.

	FREQUENCY			
POINTS TO BE INSPECTED BY	Annually / every 1400 transfer / every 50 assembling*			
USER/SERVICE TECHNICIAN	Monthly			
	Before every use	₽ <b>Ψ</b>	Ψ	Ψ
Ensure floor is clear of any dust, grease, water	or any foreign substances	•		
Make sure that all posts are straight.		•	•	•
Inspect for evidence of external damages or missing parts.			•	•
Make sure that bases are properly adjusted		•	•	•
No Red marks are visible.			•	•
Make sure track is secured in place. Press up on the track to make sure it is locked onto the post.			•	•
Clean the inside of the tracks to ensure there is no dust or grease build-up inside the tracks.			•	•
Check to make sure that the trolleys move freely along the tracks.			•	•

<sup>\*</sup> Which ever comes first: Annually - every 1400 patient transfer cycles - every 50 assembling/disassembling cycles.

WARNING: If any of the above inspection point fails, DO NOT USE the EASYTRACK FS REP system until appropriate corrective action is taken. Failure to comply with this notice could result in serious injuries.

#### **Annual Inspection by an Authorized Technician**

The EASYTRACK FS REP must be inspected by a qualified ArjoHuntleigh technician on an annual basis to ensure the proper functioning of its components, as well as the safety of the patients and care givers that use the EASYTRACK FS REP. Please contact your local representative to make arrangements.

**NOTE**: If the product does not work as intended, immediately contact your local ArjoHuntleigh agent for support.

#### Cleaning

To clean the EASYTRACK FS REP, wipe it down with a damp cloth using warm water and mild detergent.

To ensure a better rolling surface for the trolley wheels, clean the inside of the track every four months. To do so, insert a damp cloth in the opening and slide it from one end of the track to the other.

# 001.17310.EN rev. 5

14

### Troubleshooting

PROBLEM	TO CHECK
Trolley moves substantially on its own when the Easytrack FS system is fully assembled.	<ul> <li>Is the track levelled?</li> <li>Is the floor levelled?</li> <li>Is the track secured to the posts? Press up on the track to make sure that it is correctly fixed onto posts.</li> </ul>
Trolley does not roll in track smoothly.	<ul> <li>Is track clean? Clean track with damp cloth.</li> <li>Is trolley centered in track? Re-center trolley.</li> <li>Is the trolley damaged? Replace trolley</li> </ul>
The trolley does not move	<ul> <li>Did you extend the inner track in order to release it from the track slot (transportation)?</li> <li>Is there debris on the track?</li> </ul>
Extendable track comes completely apart during assembly.	<ul> <li>Track has been extended too quickly/forcefully. Track was pulled past the locking device.</li> <li>Check locking device inside track (top of track above the trolley) to make sure it is not damaged. Replace the trolley if damaged. See "Product Description / Labels" on page 5 .</li> <li>Put track back together at a slight angle and slide together.</li> <li>Slowly extend track out again to ensure locking device clicks into place.</li> <li>Slowly compress the track again.</li> </ul>
Posts and bases rock back and forth	<ul> <li>Check integrity of the floor.</li> <li>Are the levellers properly adjusted?</li> <li>Are the posts properly inserted in the base?</li> <li>Are the collar clamps closed?</li> </ul>
It is impossible to retract posts	<ul><li>Did you release the collar clamps?</li><li>Did you follow the disassembling instructions?</li><li>Did you remove lift from trolley?</li></ul>
Main hook does not engage on track pin	<ul><li>Is there debris in the track?</li><li>Is the track well seated on the posts?</li></ul>

### **Technical Specifications**

PRODUCT INFORMATION	EASYTRACK FS REP № 9720010
Weight	Total: 38.2 kg (84.2 lb) Post: 7.75 kg (17 lb) (each) Track: 13.8 kg (30.4 lb) Base: 4.45 kg (9.8 lb) (each)
Max. safe working load	200 kg (440 lb)
Range	Extendable Track Length: 2210 - 3230 mm (87 - 127") Post length: 2406 - 1830 m, (72 - 94 1/2")
Material	Aluminium
RECYCLING	
Package	Cardboard recyclable
The EASYTRACK FS REP	Separated and recycled

### **Dimensions**




Fig.25

#### **AUSTRALIA**

ArjoHuntleigh Pty Ltd 78, Forsyth street O'Connor

AU-6163 Western Australia Tel: +61 89337 4111

Free: +1 800 072 040 Fax: + 61 89337 9077

#### **BELGIQUE / BELGIË**

ArjoHuntleigh NV/SA Evenbroekveld 16 B-9420 ERPE-MERE Tél/Tel: +32 (0) 53 60 73 80 Fax: +32 (0) 53 60 73 81 E-mail: info@arjohuntleigh.be

#### **CANADA**

ArjoHuntleigh 90 Matheson Boulevard West MISSISSAUGA, ON, L5R 3R3 Tel/Tél: +1 905 238 7880

Free: +1 800 665 4831 Institutional Free: +1 800 868 0441 Home Care

Fax: +1 905 238 7881

E-mail: info.canada@arjohuntleigh.com

#### ČESKÁ REPUBLIKA

ArjoHuntleigh s.r.o. Hlinky 118 CZ-603 00 BRNO Tel: +420 549 254 252 Fax: +420 541 213 550

#### DANMARK

ArjoHuntleigh A/S Vassingerødvej 52 DK-3540 LYNGE Tel: +45 49 13 84 86 Fax: +45 49 13 84 87

E-mail: info.dk@arjohuntleigh.com

#### **DEUTSCHLAND**

ArjoHuntleigh GmbH Peter-Sander-Strasse 10 D-55252 MAINZ-KASTEL Tel: +49 (0) 6134 186 0 Fax: +49 (0) 6134 186 160 E-mail: info-de@arjohuntleigh.com

#### ΕΛΛΑΔΑ

C. Psimitis Co Ltd Dimitriou Andr. 59 GR-16121 KAISARIANI ATTIKIS Τηλ: 21 0724 36 68

Φάξ: 21 0721 55 53

#### **FSPAÑA**

ArjoHuntleigh Ibérica S.L. Ctra. de Rubí, 88 1ª planta - A1 08173 Sant Cugat del Vallés ES-BARCELONA 08173 Tel: +34 93 583 11 20

Fax: +34 93 583 11 22

E-mail: info.es@arjohuntleigh.com

#### FRANCE

ArioHuntleigh SAS 2 Avenue Alcide de Gasperi CS 70133 59436 RONCQ CEDEX Tél: +33 (0) 3 20 28 13 13

Fax: +33 (0) 3 20 28 13 14 E-mail: info.france@arjohuntleigh.com

#### HONG KONG

ArjoHuntleigh (Hong Kong) Ltd 1510-17, 15/F, Tower 2, Kowloon Commerce Centre, 51 Kwai Cheong Road, Kwai Chung, NT, HONG KONG Tel: +852 2207 6363 Fax: +852 2207 6368

#### **INTERNATIONAL**

ArjoHuntleigh International Ltd. ArjoHuntleigh House Houghton Hall Park Houghton Regis **UK-DUNSTABLE LU5 5XF** Tel: +44 (0) 1582 745 800 Fax: +44 (0) 1582 745 866

international@ArjoHuntleigh.com

#### **ITALIA**

ArjoHuntleigh S.p.A. Via di Tor Vergata 432 00133 ROMA - ITALIA Tel: +39 (0) 6 87426211 Fax: +39 (0) 6 87426222

E-mail: Italy.promo@arjohuntleigh.com

#### **NEDERLAND**

ArjoHuntleigh Nederland BV Biezenwei 21 4004 MB TIEL Postbus 6116 4000 HC TIEL

Tel: +31 (0) 344 64 08 00 Fax: +31 (0) 344 64 08 85 E-mail: info.nl@arjohuntleigh.com

#### **NEW ZEALAND**

ArjoHuntleigh Ltd 41 Vestey Drive Mount Wellington **AUCKLAND 1060** Tel: +64 (0) 9 573 5344 Free Call: 0800 000 151 Fax: +64 (0) 9 573 5384 E-mail: nz.info@ArjoHuntleigh.com

#### **NORGE**

ArjoHuntleigh Norway AS Ryenstubben 2 NO-0679 OSLO Tel: +47 22 08 00 50 Faks: +47 22 08 00 51

E-mail: no.kundeservice@arjohuntleigh.com

#### ÖSTERREICH

ArjoHuntleigh GmbH Dörrstrasse 85 AT-6020 INNSBRUCK Tel: +43 (0) 512 204 160 0 Fax: +43 (0) 512 204 160 75

#### **POLSKA**

ArjoHuntleigh Polska Sp. z o.o. ul. Ks Piotra Wawrzyniaka 2 PL 62-052 KOMORNIKI (Poznan) Tel: +48 61 662 15 50

Fax: +48 61 662 15 90

E-mail: arjo@arjohuntleigh.com

#### **PORTUGAL**

ArjoHuntleigh em Portugal: MAQUET Portugal, Lda. (Distribudor Exclusivo) Rua Poeta Bocage n.º 2 - 2G 1600-233 Lisboa, Portugal Tel: +351 214 189 815 Fax: +351 214 177 413

E-mail: Portugal@arjohuntleigh.com

#### SUISSE / SCHWEIZ

ArjoHuntleigh AG Fabrikstrasse 8 Postfach 4614 Hägendorf,

Tél/Tel: +41 (0) 61 337 97 77 Fax: +41 (0) 61 311 97 42

ArjoHuntleigh OY Vanha Porvoontie 229 FI-01380 VANTAA Puh: +358 9 4730 4320 Faksi: +358 9 4730 4999

#### **SVERIGE**

ARJO Scandinavia AB Verkstadsvägen 5 Box 61 SE-241 21 ESLÖV

Tel: +46 (0) 10-335 45 00 Fax: +46 (0) 413-138 76

E-mail: kundservice@arjohuntleigh.com

#### **UNITED KINGDOM**

ArjoHuntleigh UK ArjoHuntleigh House Houghton Hall Park Houghton Regis **UK-DUNSTABLE LU5 5XF** Tel: +44 (0) 1582 745 700

Fax: +44 (0) 1582 745 745

sales.admin@ArjoHuntleigh.com

ArjoHuntleigh Inc. 2349 W Lake Street Suite 250 Addison, IL 60101

Tel: +1 630 307 2756

Free: +1 800 323 1245 Institutional Free: +1 800 868 0441 Home Care

Fax: +1 630 307 6195

E-mail: us.info@ArjoHuntleigh.com



**GETINGE GROUP** is a leading global provider of products and systems that contribute to quality enhancement and cost efficiency within healthcare and life sciences. We operate under the three brands of **ArjoHuntleigh**, **GETINGE** and **MAQUET**. **ArjoHuntleigh** focuses on patient mobility and wound management solutions. **GETINGE** provides solutions for infection control within healthcare and contamination prevention within life sciences. **MAQUET** specializes in solutions, therapies and products for surgical interventions and intensive care.

www.ArjoHuntleigh.com



