

# **QUANTUM**

**LITESTREAM XF**



**OWNER'S MANUAL**

# SAFETY GUIDELINES



**WARNING! A Quantum Rehab Provider or a qualified technician must perform the initial setup of this wheelchair and must perform all of the procedures in this manual.**

The symbols below are used throughout this owner's manual and on the wheelchair to identify warnings and important information. It is very important for you to read them and understand them completely.



**WARNING! Indicates a potentially hazardous condition/situation. Failure to follow designated procedures can cause either personal injury, component damage, or malfunction. On the product, this icon is represented as a black symbol on a yellow triangle with a black border.**



**MANDATORY! These actions should be performed as specified. Failure to perform mandatory actions can cause personal injury and/or equipment damage. On the product, this icon is represented as a white symbol on a blue dot with a white border.**



**PROHIBITED! These actions are prohibited. These actions should not be performed at any time or in any circumstances. Performing a prohibited action can cause personal injury and/or equipment damage. On the product, this icon is represented as a black symbol with a red circle and red slash.**

## Quick Reference Information

Quantum Rehab Provider: \_\_\_\_\_

Address: \_\_\_\_\_

Phone Number: \_\_\_\_\_

Purchase Date: \_\_\_\_\_

**NOTE: This owner's manual is compiled from the latest specifications and product information available at the time of publication. We reserve the right to make changes as they become necessary. Any changes to our products may cause slight variations between the illustrations and explanations in this manual and the product you have purchased. The latest/current version of this manual is available on our website.**

**NOTE: This product is compliant with WEEE, RoHS, and REACH directives and requirements.**



Copyright © 2010

Pride Mobility Products Corp.

INFMANU2788/Rev H/May 2010

# CONTENTS

<b>I. INTRODUCTION</b> .....	4
<b>II. SAFETY</b> .....	6
<b>III. YOUR WHEELCHAIR</b> .....	15
<b>IV. COMFORT ADJUSTMENTS</b> .....	19
<b>V. CARE AND MAINTENANCE</b> .....	31
<b>VI. WARRANTY</b> .....	35
<b>APPENDIX I - SPECIFICATIONS</b> .....	36
<b>APPENDIX II - STF HEIGHT MATRIX</b> .....	39

# I. INTRODUCTION

## **SAFETY**

WELCOME to Quantum Rehab, a division of Pride Mobility Products (Pride). The wheelchair you have purchased combines state-of-the-art components with **safety**, comfort, and styling in mind. We are confident that these design features will provide you with the conveniences you expect during your daily activities. Once you understand how to **safely** operate and care for your wheelchair, it should give you years of trouble free operation and service.

**Read and follow** all instructions, warnings, and notes in this manual before attempting to operate your wheelchair for the first time. You must also read all instructions, warnings, and notes contained in any supplemental instructional booklets that accompanied your wheelchair before initial operation. Your **safety** depends upon you, as well as your provider, caretaker, or healthcare professional in using good judgement.

If there is any information in this manual which you do not understand, or if you require additional assistance for setup or operation, please contact your Quantum Rehab Provider. **Failure to follow the instructions in this manual and those located on your wheelchair can lead to personal injury and/or damage to the wheelchair, including voiding the warranty.**

## **PURCHASER'S AGREEMENT**

By accepting delivery of this product, you promise that you will not change, alter, or modify this product or remove or render inoperable or unsafe any guards, shields, or other safety features of this product; fail, refuse, or neglect to install any retrofit kits from time to time provided by Pride to enhance or preserve the safe use of this product.

## **SHIPPING AND DELIVERY**

Before using your wheelchair, make sure your delivery is complete as some components may be individually packaged. If you do not receive a complete delivery, please contact your Quantum Rehab Provider immediately. Where damage has occurred during transport, either to the packaging or content, please contact the delivery company responsible.

## **INFORMATION EXCHANGE**

We want to hear your questions, comments, and suggestions about this manual. We would also like to hear about the safety and reliability of your new wheelchair, and about the service you received from your Quantum Rehab Provider.

Please notify us of any change of address, so we can keep you apprised of important information about safety, new products, and new options that can increase your ability to use and enjoy your wheelchair. Please feel free to contact us at one of the following addresses:

### **USA:**

Pride Mobility Products Corporation  
Attn.: Customer Care Department  
182 Susquehanna Avenue  
Exeter, PA 18643-2694  
info@pridemobility.com  
1-800-424-8205

### **Canada:**

Pride Mobility Products Company  
380 Vansickle Road Unit 350  
St. Catharines, Ontario L2R 6P7  
1-888-570-1113

# I. INTRODUCTION

**United Kingdom:**

Pride Mobility Products Ltd.  
32 Wedgwood Road  
Bicester, Oxon OX26 4UL  
UK

**Australia:**

Pride Mobility Products Australia Pty. Ltd.  
21 Healey Road  
Dandenong, 3175  
Victoria, Australia

***NOTE: If you ever lose or misplace your product registration card or your copy of this manual, contact us and we will be glad to send you a new one immediately.***


## II. SAFETY

### PRODUCT SAFETY SYMBOLS

The symbols below are used on the wheelchair to identify warnings, mandatory actions, and prohibited actions. It is very important for you to read and understand them completely.



**Read and follow the information in the owner's manual.**



**Maximum seating weight.**



**Keep your hands away from the tires when driving. Be aware that loose fitting clothing can become caught in drive tires.**



**Avoid exposure to rain, snow, ice, salt, or standing water whenever possible. Maintain and store in a clean and dry condition.**



**Pinch/Crush points created during assembly.**



**Disposal and recycling - Contact your Quantum Rehab Provider for information on proper disposal and recycling of your Pride product and its packaging.**



## II. SAFETY



**Do not allow the wheelchair to roll freely on a decline.**



**Do not allow the wheelchair to roll freely on an incline.**



**Indicates that wheelchair, with similarly labeled seating system, conforms to ANSI/RESNA WC/Vol. 4, Section 19/ISO 7176-19 for transport of an occupied wheelchair in a motor vehicle.**



**Indicates wheelchair securement points.**

# II. SAFETY

## GENERAL GUIDELINES



**MANDATORY! Do not operate your new wheelchair for the first time without completely reading and understanding this owner's manual.**

Your wheelchair is a state-of-the-art life-enhancement device designed to increase mobility. Pride provides an extensive variety of products to best fit the individual needs of the wheelchair user. Please be aware that the final selection and purchasing decision regarding the type of wheelchair to be used is the responsibility of the wheelchair user, who is capable of making such a decision, and his/her healthcare professional (i.e., medical doctor, physical therapist, etc.).

The contents of this manual are based on the expectation that a mobility device expert has properly fitted the wheelchair to the user and has assisted the prescribing healthcare professional and/or the Quantum Rehab Provider in the instruction process for the use of the product.

There are certain situations, including some medical conditions, where the wheelchair user will need to practice operating the wheelchair in the presence of a trained attendant. A trained attendant can be defined as a family member or care professional specially trained in assisting a wheelchair user in various daily living activities.

As you begin using your wheelchair during daily activities, you will probably encounter situations in which you will need some practice. Simply take your time and you will soon be in full and confident control as you maneuver through doorways, on and off of elevators, up and down ramps, and over moderate terrain.

Below are some precautions, tips, and other safety considerations that will help you become accustomed to operating your wheelchair safely.

### Modifications

Pride has designed and engineered your wheelchair to provide maximum mobility and utility. A wide range of accessories is available from your Quantum Rehab Provider to further customize your wheelchair to better suit your needs and/or preferences. However, under no circumstances should you modify, add, remove, or disable any feature, part, or function of your wheelchair.



**WARNING! Do not modify your wheelchair in any way not authorized by Pride. Do not use accessories if they have not been tested or approved for Pride products.**

### Pre-Ride Safety Check

Get to know the feel of your wheelchair and its capabilities. Pride recommends that you perform a safety check before each use to make sure your wheelchair operates smoothly and safely. Before riding, you should be trained in the safe use of this wheelchair by your Quantum Rehab Provider.

### Perform the following inspections prior to using your wheelchair:

- Check for proper tire inflation. Maintain but do not exceed the psi/bar/kPa air pressure rating indicated on each tire if equipped with pneumatic tires.
- Check the brakes. See V. "Care and Maintenance."

**NOTE: If you discover a problem, contact your Quantum Rehab Provider for assistance.**

## II. SAFETY

### Weight Limitations

Your wheelchair is rated for a maximum weight capacity. Please refer to the specifications table for this limit. Keep in mind that the maximum weight capacity includes the combined weight of the user and any accessories mounted to the wheelchair.



**MANDATORY!** Stay within the specified weight capacity of your wheelchair. Exceeding the weight capacity voids your warranty. Pride will not be held responsible for injuries and/or property damage resulting from failure to observe weight limitations.

**WARNING!** Do not carry passengers on your wheelchair. Carrying passengers on your wheelchair may affect the center of gravity, resulting in a tip or a fall.



**WARNING!** The rear flap is for pocket items such as cell phones, keys, and papers. Avoid storing sharp objects in the rear flap. There are no other storage areas on your wheelchair.

**WARNING!** Do not carry heavy baggage weighing more than 10 lbs. (4.5 kg) on the back of the wheelchair. Carrying or tying baggage in excess of 10 lbs. (4.5 kg) on the back of the wheelchair may affect the center of gravity, resulting in a tip or a fall.

### Tire Inflation

If your wheelchair is equipped with pneumatic tires, you should check or have the air pressure checked regularly. Proper inflation pressures will prolong the life of your tires and help ensure the smooth operation of your wheelchair.



**WARNING!** It is important that the psi/bar/kPa air pressure rating indicated on the tire be maintained in pneumatic tires at all times. Do not underinflate or overinflate your tires. Low pressure may result in loss of control, and overinflated tires may burst. Failure to maintain the psi/bar/kPa air pressure rating indicated on the pneumatic tires at all times may result in tire and/or wheel failure.

**WARNING!** Inflate your wheelchair drive tires from a regulated air source with an available pressure gauge. Inflating your tires from an unregulated air source could overinflate them, resulting in a burst tire.

**NOTE:** If the tire on your wheelchair lists the psi rating only, use the following conversion formulas to find the bar or kPa rating:  $bar = psi \times 0.06895$ ;  $kPa = psi \times 6.89476$ .

### Incline Information

More and more buildings have ramps with specified degrees of inclination, designed for easy and safe access. Some ramps may have turning switchbacks (180-degree turns) that require you to have good cornering skills on your wheelchair.

- Proceed with extreme caution as you approach the downgrade of a ramp or other incline.
- Take wide swings with your wheelchair's front wheels around any tight corners. If you do that, the wheelchair's rear wheels will follow a wide arc, not cut the corner short, and not bump into or get hung up on any railing corners.
- When driving down a ramp, proceed slowly to ensure a safely controlled descent.
- Avoid sudden stops and starts.

When climbing an incline, try to keep your wheelchair moving. If you must stop, start up again slowly and then accelerate cautiously. When driving down an incline, proceed slowly and drive in the forward direction only. If your wheelchair starts to move down the incline faster than you anticipated or desired, bring it to a complete stop using both hands, then proceed slowly to ensure a safely controlled descent.



**WARNING!** When on any sort of an incline or decline, never allow the wheelchair to roll freely while seated on it or standing next to it.

**WARNING!** When climbing an incline, do not zigzag or drive at an angle up the face of the incline. Drive your wheelchair straight up the incline. This greatly reduces the possibility of a tip or a fall. Always exercise extreme caution when negotiating an incline.



**WARNING!** You should not travel up or down a potentially hazardous incline (i.e., areas covered with snow, ice, cut grass, or wet leaves).

**WARNING!** Never travel down an incline backward. Doing so may cause the wheelchair to tip. Always exercise extreme caution when negotiating an incline.

## II. SAFETY

**WARNING! If your wheelchair is equipped with a reclining seatback, do not attempt to negotiate inclines with the seat in a reclined position. Do not attempt to negotiate obstacles with the seat in a reclined position unless an attendant is present to help stabilize the chair. Failure to heed may result in the wheelchair tipping over.**



**WARNING! If your wheelchair is equipped with an optional carry bag mounted to the seatback, do not overload the bag or exceed the maximum load rating of the bag. Doing so could cause instability in your wheelchair when negotiating inclines, causing it to tip.**

**WARNING! Even though your wheelchair is capable of climbing slopes greater than those illustrated in figure 1, do not, under any circumstances, exceed the incline guidelines or any other specifications presented in this manual. Doing so could cause instability in your wheelchair.**

Pride recommends that the maximum slope of an incline you attempt to safely ascend or descend on your wheelchair does not exceed 5° (8.7%). See figure 1.



**WARNING! Any attempt to climb or descend a slope steeper than 5° (8.7%) may put your wheelchair in an unstable position and cause it to tip.**




Figure 1. Maximum Safe Slope (Ascending and Descending)

### Braking Information

To stop the wheelchair, apply pressure to both handrims until the wheelchair comes to a complete stop.

### Manual Wheel Locks

Your wheelchair is equipped with two manual wheel locks; one mounted near (or directly in front of) each drive wheel. The manual wheel locks help prevent your wheelchair from moving when parked. After you bring your wheelchair to a complete stop, engage both manual wheel locks. There are three types of manual wheel locks available on your wheelchair: push-to-lock, pull-to-lock, and scissor-lock. If you have any questions about the operation of your manual wheel locks, contact your Quantum Rehab Provider.



**WARNING! Do not attempt to use either manual wheel lock lever to stop your wheelchair.**



**WARNING! Pride does not recommend that you remain parked on an incline for a long period of time.**

**WARNING! Use care when operating the wheel locks. Possible pinch point created during operation.**

### Cornering Information

While your wheelchair is equipped with front caster wheels, excessively high cornering speeds can still create the possibility of tipping. Factors which affect the possibility of tipping include, but are not limited to: cornering speed, steering angle (how sharply you are turning), uneven road surfaces, inclined road surfaces, riding from an area of low traction to an area of high traction (such as passing from a grassy area to a paved area – especially at high speed while turning), and abrupt directional changes. High cornering speeds are not recommended. If you feel that you may tip over in a corner, reduce your speed and steering angle (i.e., lessen the sharpness of the turn) to prevent your wheelchair from tipping.



**WARNING! When cornering sharply, reduce your speed and maintain a stable center of gravity. This greatly reduces the possibility of a tip or fall.**

### Public Streets and Roadways



**WARNING! You should not operate your wheelchair on public streets and roadways. Be aware that it may be difficult for traffic to see you when you are seated on your wheelchair. Obey all local pedestrian traffic rules. Wait until your path is clear of traffic, and then proceed with extreme caution.**

## II. SAFETY

### Outdoor Driving Surfaces

Your wheelchair is designed to provide optimum stability under normal driving conditions—dry, level surfaces composed of concrete, blacktop, or asphalt. However, Pride recognizes that there will be times when you will encounter other surface types. For this reason, your wheelchair is designed to perform admirably on packed soil, grass, and gravel. Feel free to use your wheelchair safely on lawns and in park areas.

- Avoid driving on uneven terrain and/or soft surfaces.
- Avoid tall grass that can entangle the drive wheels or caster wheels.
- Avoid loosely packed gravel and sand.
- Avoid uneven surfaces.
- Do not use your wheelchair in fresh or salt water.
- Do not use your wheelchair at the edge of a stream, lake, or ocean.
- If you feel unsure about a driving surface, avoid that surface.

### Inclement Weather Precautions

Exposure of your wheelchair to inclement weather conditions should be avoided whenever possible. If suddenly caught up in rain, snow, severe cold or heat while operating your wheelchair, proceed to shelter at the earliest opportunity. Thoroughly dry your wheelchair before storing or operating your wheelchair.



**PROHIBITED!** Operating in rain, snow, salt, mist/spray conditions, and on icy/slippery surfaces can have an adverse affect on your wheelchair. Maintain and store your wheelchair in a dry and clean condition.



**WARNING!** Prolonged exposure to hot or cold conditions may affect the temperature of upholstered and non-upholstered items on the wheelchair, possibly resulting in skin irritation. Exercise caution when using the wheelchair in extremely hot or cold conditions or when exposing your wheelchair to direct sunlight for prolonged periods of time.

### Stationary Obstacles (Steps, Curbs, etc.)

Proceed with extreme caution when driving near raised surfaces, unprotected ledges and/or drop-offs (curbs, porches, stairs, etc.). Be sure your wheelchair is traveling perpendicular to any curb you may be required to navigate. See figure 2.



**WARNING!** Even though your wheelchair may be capable of handling greater obstacles, Pride recommends that you do not attempt to negotiate a curb that has a height greater than 2 in. (5 cm). Doing so could cause instability in your wheelchair.

**WARNING!** Do not attempt to have your wheelchair proceed backward down any step, curb, or other obstacle. This may cause the wheelchair to tip.




Figure 2. Correct Curb Approach



Figure 3. Incorrect Curb Approach

### Stairs and Escalators

Wheelchairs are not designed to travel up or down stairs or escalators. Always use an elevator.



**WARNING!** Never use your wheelchair to negotiate steps or escalators.

## II. SAFETY

### Doors

- Determine if the door opens toward or away from you.
- Drive your wheelchair gently and slowly forward to push the door open. Or drive your wheelchair gently and slowly backward to pull the door open.

### Elevators

Modern elevators have a door edge safety mechanism that, when pushed, reopens the elevator door(s).

- If you are in the doorway of an elevator when the door(s) begin to close, push on the rubber door edge or allow the rubber door edge to contact the wheelchair and the door will reopen.
- Use care that pocketbooks, packages, or wheelchair accessories do not become caught in elevator doors.

### Lift Systems/Elevation Products

If you will be transporting your wheelchair during travel, you may find it necessary to use a vehicle-mounted lift system or elevation product to aid in transportation. Pride recommends that you closely review the warnings, instructions, specifications, and safety information set forth by the manufacturer of the lift/elevation product before using that product.



**WARNING! Not all lift systems/elevation products are rated for use with an occupied mobility device. Pride recommends that you do not remain seated in your wheelchair when it is being used in connection with any type of lift system/elevation product unless:**

- That product is rated for occupied use by the manufacturer.
- You are able to occupy your mobility device in strict conformity with the warnings, instructions, specifications, and safety guidelines set forth by the manufacturer.

### Motor Vehicle Transport

Pride recommends that you do not remain seated in your wheelchair while traveling in a motor vehicle. Your wheelchair was designed to be folded and stored in a secure location, such as the trunk of a car, during motor vehicle transport, unless it is equipped with an optional transit securement system. If your wheelchair is equipped with a transit securement system, refer to the Transit Securement System supplement provided with your wheelchair.



**WARNING! The wheelchair user should transfer into the motor vehicle seat and use the vehicle-installed restraint system whenever feasible. The wheelchair user should never sit in the wheelchair while it is in a moving vehicle unless it is equipped and properly secured with a manufacturer-installed transit securement system.**

**WARNING! Ensure wheelchairs equipped with a transit securement system are properly secured to the motor vehicle during transport. Wheelchairs that are not properly secured can become a hazard to the user and to other vehicle passengers in the event of a crash, sudden stopping, or swerving, as the wheelchair could tip or slide out of place.**

### Positioning Belts

Your Quantum Rehab Provider, therapist(s), and other healthcare professionals are responsible for determining your requirement for a positioning belt in order to operate your wheelchair safely. Typically, the positioning belt is not designed for use as a restraining device in a motor vehicle. You should be properly belted into seats approved by the vehicle manufacturer with your wheelchair properly stowed for vehicle transport or secured using an optional transit securement system.

Positioning belts should be installed as detailed in this manual. See IV. "Comfort Adjustments." Once installed, the belt should be placed across the pelvis as determined by your healthcare provider. Anchor points of the positioning belts should be located so that a projected side-view angle of the pelvis is between 45°-75° of horizontal depending on the user's weight, posture, and seat cushion thickness.



**WARNING! If you require a positioning belt to safely operate your wheelchair, make sure it is fastened securely in order to reduce the possibility of a fall from the wheelchair.**

## II. SAFETY



**WARNING!** The positioning belt is not designed for use as a seat belt in a motor vehicle. Nor is your wheelchair suitable for use as a seat in any vehicle. Anyone traveling in a vehicle should be properly belted into seats approved by the vehicle manufacturer.

### Removable Parts



**WARNING!** Do not attempt to lift or move a wheelchair by any of its removable parts, including the armrests, seat, and front riggings. Do not attempt to pick up the wheelchair while the seat is occupied.

### Reaching and Bending

Never reach, lean, or bend while driving your wheelchair. If it is absolutely necessary to reach, lean, or bend while seated on your wheelchair, it is important to maintain a stable center of gravity and keep the wheelchair from tipping. Pride recommends that the wheelchair user determine his/her personal limitations and practice bending and reaching in the presence of a qualified healthcare professional.



**WARNING!** Do not bend, lean, or reach for objects if you have to pick them up from the floor by reaching down between your knees or if you have to reach over the back of the seat. Movements such as these may change your center of gravity and the weight distribution of the wheelchair. This may cause your wheelchair to tip.



**WARNING!** Keep your hands away from the tires when driving. Be aware that loose fitting clothing can become caught in drive tires.

### Transfers

Transferring onto and off of your wheelchair requires a good sense of balance. Always have an attendant or healthcare professional present while learning to properly transfer yourself.

To eliminate the possibility of injury, Pride recommends that you or a trained attendant perform the following tasks before attempting a transfer:

- Engage the caster pin lock, if equipped. See III. "Your Wheelchair."

**NOTE:** If your wheelchair is equipped with a caster pin lock, be sure to disengage the caster pin lock before operating your wheelchair. See III. "Your Wheelchair."

- If your wheelchair is not equipped with a caster pin lock, turn both caster wheels toward the transfer destination to improve wheelchair stability during transfer.
- Ensure both wheel locks on your wheelchair are engaged. See III. "Your Wheelchair."
- If necessary, remove or flip up one or both armrests. See IV. "Comfort Adjustments."
- Move the leg rests aside; this will help to keep your feet from getting caught on the front rigging during the transfer.
- Reduce the distance between your wheelchair and the object you are transferring onto.




Figure 4. Ideal Transfer Position

## II. SAFETY

**WARNING! Before transferring, position yourself as far back as possible in the wheelchair seat to prevent the wheelchair from tipping forward during transfer.**

**WARNING! Avoid putting all of your weight on the wheelchair armrests and do not use the armrests for weight bearing purposes, such as transfers. Such use may cause the wheelchair to tip, resulting in a fall from the wheelchair and personal injury.**



**WARNING! Avoid putting all of your weight on the manual wheel locks and do not use the manual wheel locks for weight bearing purposes, such as transfers. Such use may cause equipment damage or personal injury.**

**WARNING! Avoid putting all of your weight on the front riggings. Such use may cause the wheelchair to tip.**

**WARNING! Ensure legs, hands, and clothing are clear of the scissor-style wheel lock levers when attempting to transfer.**

### Prescription Drugs/Physical Limitations

Users must exercise care and common sense when operating a wheelchair. This includes awareness of safety issues when taking prescribed or over-the-counter drugs or when the user has specific physical limitations.



**WARNING! Consult your physician if you are taking prescribed or over-the-counter medication or if you have certain physical limitations. Some medications and limitations may impair your ability to operate your wheelchair in a safe manner.**

### Alcohol/Smoking

The wheelchair user must exercise care and common sense when operating his/her wheelchair. This includes awareness of safety issues while under the influence of alcohol or while smoking.

**WARNING! Do not operate your wheelchair while you are under the influence of alcohol, as this may impair your ability to operate your wheelchair in a safe manner.**



**WARNING! Pride strongly recommends that you do not smoke cigarettes while seated in your wheelchair, although the wheelchair has passed the necessary testing requirements for cigarette smoking. You must adhere to the following safety guidelines if you decide to smoke cigarettes while seated in your wheelchair.**

- Do not leave lit cigarettes unattended.
- Keep ashtrays a safe distance from the seat cushions.
- Always make sure cigarettes are completely extinguished before disposal.

# III. YOUR WHEELCHAIR

## THE LITESTREAM XF

The Litestream XF is a dual-purpose, rear-wheel drive wheelchair designed to operate both indoors and outdoors on smooth surfaces.




Figure 5. The Litestream XF Wheelchair

# III. YOUR WHEELCHAIR

## USING THE LITESTREAM XF

The Litestream XF is a manual wheelchair. Either an attendant pushes it, or you move it by gripping the hand-rims (see figure 5) and turning them in the desired direction of travel. You steer it by turning one wheel faster than the other. To stop the wheelchair, you apply pressure to both handrims until the wheelchair comes to a complete stop.

## Unfolding

The Litestream XF is a folding wheelchair. Unfold it completely before sitting in it.

### To unfold the Litestream XF:

1. Engage both wheel locks. See figure 6.



**WARNING!** Use care when operating the manual wheel locks. Possible pinch point created during operation.

2. Place the palms of your hands on each seat base rail. See figure 7.
3. Press down until the seat base rails are locked into place.

**NOTE:** If your wheelchair is equipped with an HD option, you need to secure the stabilizing bar prior to use.

### To secure the stabilizing bar:

1. Press and hold the release mechanism toward the center of the wheelchair. See figure 8.
2. Push the stabilizing bar down onto the back cane.
3. Release the mechanism.
4. Ensure the stabilizing bar is locked into place.

## Folding

Folding the wheelchair makes it easier to carry and store. The folded width is 11.75 in. (29.85 cm).

### To fold the Litestream XF:

1. Engage both wheel locks. See figure 6.
2. Flip up the foot plates (if applicable).
3. If your wheelchair is equipped with an HD option, unlatch the stabilizing bar by pressing and holding the release mechanism and lifting the stabilizing bar up and away from the wheelchair.
4. Pull up on the strap until both wheels close together. See figure 9.




Figure 6. Right Wheel Lock Engaged (Push-to-lock Shown)




Figure 7. Unfolding the Litestream XF




Figure 8. Stabilizing Bar Operation

# III. YOUR WHEELCHAIR

**NOTE:** If your wheelchair is equipped with a depth-adjustable side frame, height-adjustable armrests, and waterfall armrests, interference may be encountered when attempting to fold the wheelchair. It may be necessary to remove the drive wheels to fold the wheelchair.

## Transporting Your Litestream XF

The Litestream XF was designed to be folded and stored in a secure location, such as the trunk of a car, during motor vehicle transport, unless it is equipped with an optional transit securement system. If your wheelchair is equipped with a transit securement system, refer to the Transit Securement System supplement provided with your wheelchair.

### To stow the Litestream XF:

1. If your wheelchair is equipped with an HD option, unlatch the stabilizing bar by pressing and holding the release mechanism and lifting the stabilizing bar up and away from the wheelchair. See figure 8.
2. Fold the Litestream XF.
3. With one hand, grab and squeeze together both back cane handles. See figure 10.
4. With the other hand, grab and squeeze together both seat base rails.

**NOTE:** You can also remove the drive wheels to make the wheelchair more compact. See “Drive Wheel Adjustments.”



**WARNING!** If you remove the drive wheels, make sure the quick-release mechanism snaps into place when reinstalling the wheel. See “Drive Wheel Adjustments.”

### Folding Down the Seatback of your Litestream XF

If your Litestream XF is equipped with fold-forward back canes, the back canes can be folded down for easier transport.

### To fold down the back canes on the Litestream XF:

1. Remove the drive wheels. See figure 15.
2. Flip up the armrests. If your wheelchair is equipped with latching armrests, push the armrest latch toward the front of the armrest to disengage the armrest. You will hear a click when the armrest is latched into place. See figure 11.
3. Pull the strap at the rear of the seat base away from the wheelchair to fold down the seatback. See figure 12.




Figure 9. Folding the Litestream XF




Figure 10. Lifting Points




Figure 11. Armrest Latch - Latching Armrests

# III. YOUR WHEELCHAIR

## Optional Caster Pin Lock

If your wheelchair is equipped with a caster pin lock, you can lock the caster wheels in the forward position to aid in transferring. The caster pin lock should be engaged prior to transfer and disengaged prior to operating the wheelchair.

### To operate the caster pin lock:

1. Rotate the lever so that the securement pin is not in the groove to engage the caster pin lock. See **figure 14**.
2. Rotate the lever so that the securement pin is securely in the groove to disengage the caster pin lock. See **figure 14**.

## Optional Transit Securement System

If your wheelchair is equipped with a transit securement system, it is equipped with manufacturer-installed front and rear securement brackets and positioning belt anchoring brackets. See **figure 13**. Refer to the Transit Securement System supplement provided with your wheelchair for information on how to use your transit securement system.

**NOTE:** The transit securement system is not available on wheelchairs equipped with an HD option.

**WARNING!** If your wheelchair is equipped with a transit securement system, the drive wheel height, seatback height, and caster angle should only be adjusted by a Quantum Rehab Provider or a qualified technician.



**WARNING!** Always use the manufacturer-installed front and rear securement brackets supplied with the transit securement system. Do not modify the transit system securement brackets in any way. Do not modify your wheelchair in any way unless authorized by Pride.




Figure 12. Folding the Seatback Down




Figure 13. Securement Brackets




Figure 14. Caster Pin Lock Operation

# IV. COMFORT ADJUSTMENTS

## COMFORT ADJUSTMENTS

After becoming familiar with your wheelchair's operation, you may find the need to make some adjustments to increase your comfort, such as drive wheel position, seatback depth, anti-tip wheels position, seatback angle and height, swing-away footrest length, elevating leg rest length and angle, calf pad height and depth, angle adjustable foot plate angle and depth, armrest height and angle, seat angle, caster angle, positioning belt, wheel lock position, and seat tension.



**WARNING!** The center of gravity of your wheelchair was factory set to a position that meets the needs of the demographic majority of users. Your Quantum Rehab Provider has evaluated your wheelchair and made any necessary adjustments to suit your specific requirements. Do not change your seating configuration without first contacting Pride Mobility Products or your Quantum Rehab Provider.

**WARNING!** Some wheelchair components are heavy. You may need assistance to lift or carry them. Please refer to "Appendix I - Specifications" for specific component weights before you disassemble the wheelchair.

**WARNING!** Remove the occupant from the wheelchair before making any adjustments.

You may need the following to make comfort adjustments:

- 1/8-in. wrench
- 3/8-in. wrench
- 7/16-in. wrench
- 1/2-in. wrench
- 3/4-in. wrench
- 4-mm hex key
- 5-mm hex key
- 3/16-in. hex key

**NOTE:** Any nylon insert lock nut removed during the disassembly or adjustment of the wheelchair must be replaced with a new nut. Nylon insert lock nuts should not be reused as it may cause damage to the nylon insert, resulting in a less secure fit. Replacement nylon insert lock nuts are available at local hardware stores or through your Quantum Rehab Provider.

### Drive Wheel Adjustments

Each drive wheel is attached to the frame by an axle bracket. The location of the drive wheel center bolt and nut on the axle bracket determines the drive wheel position. The location of the axle bracket on the wheelchair frame determines the drive wheel height and can affect performance and center of gravity.



**WARNING!** The more forward that you adjust your drive wheel, the more likely your wheelchair will tip over backward. Always make adjustments in small increments, and check the stability of your wheelchair with a Quantum Rehab Provider or with a spotter to prevent a tip over. We recommend that you use anti-tip tubes until you adapt to the change and are sure that you are not at risk to tip over.




Figure 15. Drive Wheel Quick Release

## IV. COMFORT ADJUSTMENTS

**NOTE:** If you change the drive wheel position, you may have to change the caster height and adjust the wheel lock. Consult your Quantum Rehab Provider.

### To change the drive wheel position:

1. Press the quick-release button and pull the drive wheel off. See figure 15.
2. Remove the center bolt and nut from the axle bracket. See figure 16.
3. Reposition the center bolt and nut forward or back to the desired position.
4. Tighten the center bolt and nut.
5. Reinstall the drive wheel. Make sure that the wheel axle is snapped securely into the axle bracket.

**NOTE:** Make sure that both drive wheels are adjusted to the same position.



**WARNING!** If your wheelchair is equipped with a transit securement system, the drive wheel height should only be adjusted by a Quantum Rehab Provider or a qualified technician.

### To change the drive wheel height:

1. Press the quick-release button and pull the drive wheel off. See figure 15.
2. Remove the hardware that fastens the axle bracket to the frame. See figure 17.
3. Reposition the axle bracket up or down to the desired position.

**NOTE:** The drive wheel height can be adjusted to 4 different positions.

4. Reinstall the hardware that fastens the axle bracket to the frame.
5. Reinstall the drive wheel. Make sure that the wheel axle is snapped securely into the axle bracket.

**NOTE:** Make sure that both drive wheels are adjusted to the same position.



**WARNING!** Ensure the quick-release axles are properly secured.




Figure 16. Drive Wheel Position Adjustment




Figure 17. Drive Wheel Height Adjustment

# IV. COMFORT ADJUSTMENTS

## Depth Adjustment

Your Litestream XF may be equipped with a depth-adjustable side frame. The depth adjustment range is 4 in. (10.16 cm) in 1-in (2.54-cm) increments.

**NOTE:** *The center of gravity of your wheelchair was factory set to a position that meets the needs of the demographic majority of users. It is the responsibility of your Quantum Rehab Provider to evaluate the wheelchair and make any adjustments to suit your specific requirements to prevent the wheelchair from tipping forward, rearward, or to the side.*

**WARNING!** Center of gravity adjustment may be necessary for optimal performance and stability. Do not perform these adjustments yourself. Wheelchair depth adjustment and center of gravity adjustment should only be performed by your Quantum Rehab Provider or a qualified technician.



To adjust the depth of the Litestream XF:

1. Remove the depth adjustment screws and angle adjustment screws that attach the back cane to the side frame on both sides of the wheelchair. See **figure 18**.
2. Position the back canes so that the desired depth is achieved.
3. Reinstall the depth adjustment screws and angle adjustment screws to secure the back canes to the side frame on both sides of the wheelchair.

## Anti-tip Wheel Adjustment

If your wheelchair is equipped with a depth-adjustable side frame, it will also be equipped with anti-tip wheels. Anti-tip wheels are an integral and important safety feature of your wheelchair. They are factory set to between 0.75-2 in. (2-5 cm), but may require adjustment to accommodate depth adjustment and center of gravity.



**PROHIBITED!** Do not remove the anti-tip wheels or modify the wheelchair in any way that is not authorized by Pride.



**WARNING!** The anti-tip wheels should only be adjusted by a Quantum Rehab Provider or a qualified technician.




Figure 18. Side Frame Depth Adjustment




Figure 19. Anti-tip Wheel Adjustment

## IV. COMFORT ADJUSTMENTS

### To adjust the anti-tip wheels:

1. Push the button on both sides of the anti-tip wheel at the same time and move the anti-tip wheel up or down to unlock it. **See figure 20.**
2. Adjust the anti-tip wheel to the desired position while aligning the anti-tip wheel with the adjustment holes.
3. Ensure that both buttons lock into place. You will hear a click when the anti-tip wheel is locked into place.
4. Repeat for the other side.

### Seatback Angle Adjustment

Your Litestream XF may be equipped with angle-adjustable back canes. If your wheelchair is equipped with a depth-adjustable side frame, the seatback angle can be adjusted to 80-120° in 5° increments. Otherwise, the seatback angle can be adjusted to 85-105° in 5° increments.

### To adjust the seatback angle of the Litestream XF:

1. Remove the drive wheel. **See figure 15.**
2. Remove the angle adjustment screw and loosen the depth adjustment screw. **See figure 20.**
3. Position the back canes at the desired angle.
4. Reinstall the angle adjustment screw and tighten the depth adjustment screw. **See figure 20.**
5. Repeat on the other side of the wheelchair.

### Seatback Height Adjustment

You can adjust the seatback to five different heights ranging between 12-16 in. (30.48-40.64 cm) or 15-19 in. (38.1-48.26 cm), depending on the back cane size. The adjustment range is 4 in. (10.16 cm) in 1-in. (2.54-cm) increments. **See figure 21.**



**WARNING!** If your wheelchair is equipped with a transit securement system, the seatback height should only be adjusted by a Quantum Rehab Provider or a qualified technician.

### To adjust the seatback height:

1. Remove the screws, nuts, and washers that fasten the back canes to each side of the frame.
2. Slide the back canes up or down to the desired height.
3. Use the screws, nuts, and washers to refasten the back canes to the frame.




Figure 20. Seatback Angle Adjustment (Depth-adjustable Side Frame Shown)




Figure 21. Seatback Height Adjustment

## IV. COMFORT ADJUSTMENTS

### Swing-Away Footrest (SFR) Removal

The swing-away footrests (SFR) can be moved or removed to aid in transfers or transporting your wheelchair.

#### To move the SFR to the side:

1. Push the release lever back. See figure 22.
2. Move the SFR aside.
3. Repeat for the other side if necessary.

#### To remove the SFR:

1. Move the SFR to the side.

**NOTE:** If your wheelchair is equipped with the ultra hemi option, the SFR must be facing forward and not swung to the side for removal.

2. Pull the SFR up. See figure 23.
3. Repeat for the other side if necessary.

### Swing-Away Footrest (SFR) Foot Plate Length

You can change the SFR foot plate length between 13.25-22 in. (33.66-55.88 cm) in 1-in. (2.54-cm) increments. This length is measured from the SFR foot plate to the top of the SFR. See figure 24.

#### To change the SFR foot plate length:

1. Remove the screws from the SFR.
2. Adjust the foot plate to the desired length.
3. Reinstall the screws into the SFR.
4. Repeat for the other side.

### Elevating Leg Rest (ELR) Foot Plate Length

You can change the ELR foot plate length between 13.25-22 in. (33.66-55.88 cm). This length is measured from the ELR foot plate to the top of the ELR. See figure 25.

#### To change the ELR foot plate length:

1. Loosen the screws on the back of the ELR.
2. Adjust the foot plate to the desired length.
3. Tighten the screws on the back of the ELR.
4. Repeat the procedure for the other side.




Figure 22. SFR Release




Figure 23. SFR Removal




Figure 24. SFR Foot Plate Adjustment

# IV. COMFORT ADJUSTMENTS

## Elevating Leg Rest (ELR) Angle

You can change the ELR angle. See figure 26.

### To change the ELR angle:

1. To raise, pull the ELR up to the desired position.
2. To lower, pull the ELR release lever and lower the ELR to the desired position.



**WARNING! Use care when operating. Possible pinch point created during operation.**

## Elevating Leg Rest (ELR) Calf Pad

You can change the height and depth of the ELR calf pad. See figures 27 and 28.

### To change the ELR calf pad height:

1. Remove the screw to unfasten the calf pad from the ELR.
2. Move the calf pad up or down to the desired position.
3. Replace the screw to refasten the calf pad to the ELR.
4. Repeat for the other side.

### To change the ELR calf pad depth:

1. Remove the screw to unfasten the calf pad from the calf pad bracket.
2. Move the calf pad in or out to the desired position.
3. Replace the screw to refasten the calf pad to the calf pad bracket.
4. Repeat for the other side.




Figure 25. ELR Foot Plate Adjustment




Figure 26. ELR Angle Adjustment




Figure 27. ELR Calf Pad Height Adjustment




Figure 28. ELR Calf Pad Depth Adjustment

## IV. COMFORT ADJUSTMENTS

### Angle Adjustable Foot Plates

The angle adjustable foot plates have two adjustments: angle and depth. The depth can be adjusted to four (4) positions in 1-in. (2.5 cm) increments. If the position of the foot plate adapter is reversed, an additional 0.75 in. (2 cm) of depth can be achieved at the most forward and most rearward positions. See **figure 29**. The foot plates can also be flipped up for ease of transfer. See **figure 29**.

#### To adjust the foot plate angle:

1. Loosen the screws in the foot plate.
2. Tilt the foot plate forward or back to the desired angle. See **figure 30**.
3. Tighten the screws to secure the foot plate in place.

#### To adjust the foot plate depth:

1. Remove the screws and nuts from the foot plate assembly. See **figure 31**.
2. Remove the foot plate and foot plate adapter from the leg rest extension.
3. Reposition the foot plate on the foot plate adapter to achieve the desired depth. See **figure 31**.
4. Reinstall the screws to the foot plate and secure with the nuts. Do not fully tighten.
5. Slide the foot plate assembly onto the leg rest extension.
6. Adjust the foot plate to the desired angle and fully tighten the screws.




Figure 29. Foot Plate Flip Up Adjustment




Figure 30. Foot Plate Angle Adjustment




Figure 31. Foot Plate Depth Adjustment

# IV. COMFORT ADJUSTMENTS

## Drop-in Armrest Removal

The drop-in armrests can be removed to aid in transfers or transporting your wheelchair.

### To remove the armrest:

1. Squeeze the armrest latch. See figure 32.
2. Pull the armrest up.
3. Repeat for the other side if necessary.



**WARNING! Use care when operating. Possible pinch point created during operation.**

## Drop-in Armrest Height Adjustment

You can adjust the drop-in armrest height to five positions in 1-in. (2.54-cm) increments.

### To adjust the drop-in armrest:

1. Pull back the release lever located on the armrest. See figure 33.
2. Move the armrest up or down.
3. Release the release lever.
4. Make sure the armrest latches into place.
5. Repeat for the other side if necessary.

## Flip-up Armrest Height Adjustment

You can adjust both the latching and non-latching flip-up armrest height to eight positions in 1-in. (2.54-cm) increments.

### To adjust the flip-up armrest height:

1. Remove the armrest height adjustment screws. See figure 34.
2. Move the armrest up or down.
3. Reinstall the armrest height adjustment screws.
4. Repeat for the other side if necessary.




Figure 32. Drop-in Armrest Removal




Figure 33. Drop-in Armrest Height Adjustment




Figure 34. Flip-up Armrest Height Adjustment

## IV. COMFORT ADJUSTMENTS

### Flip-up Armrest Angle Adjustment

You can adjust both the latching and non-latching flip-up armrest angle.

#### To adjust the non-latching flip-up armrest angle:

1. Flip up the armrest.
2. Turn the adjusting screw to raise or lower the front of the armrest. See figure 35.
3. Repeat for the other side if necessary.

#### To adjust the latching flip-up armrest angle:

1. Push the armrest latch toward the front of the armrest and flip up the armrest until you hear a click. See figure 11.
2. Remove the adjusting screw. See figure 36.
3. Reinstall the adjusting screw in the desired position.
4. Repeat for the other side if necessary.

### Dual-post Armrests

The dual-post armrests can be removed to aid in transfers or transporting your wheelchair.

#### To flip the dual-post armrest back:

1. Pull back the armrest latch located on the armrest. See figure 37.
2. Flip the armrest back.
3. Repeat the procedure for the other side if necessary.

#### To remove the dual-post armrest:

1. Pull back the armrest latch located on the armrest. See figure 37.
2. Flip the armrest back.
3. Push the top release button.
4. Lift the armrest off of the wheelchair
5. Repeat the procedure for the other side if necessary.




Figure 35. Non-latching Flip-up Armrest Angle Adjustment




Figure 36. Locking Flip-up Armrest Angle Adjustment




Figure 37. Dual-post Armrest Removal

## IV. COMFORT ADJUSTMENTS

### To adjust the dual-post armrest height:

1. Rotate the release lever that is located on the armrest in either direction. See **figure 38**.
2. Move the armrest up or down to the desired height.
3. The release lever is spring-loaded and will automatically release.
4. Make sure the armrest latches into place.
5. Repeat the procedure for the other side if necessary.

### Seat Angle (Dump) Adjustment

The position of the caster wheels affects the angle (dump) of the seat. By changing the location of the caster wheel on the fork, you change the seat angle. See **figure 39**.

**NOTE:** You can also adjust the seat angle by raising or lowering the drive wheel. See “*Drive Wheel Adjustment.*”

### To change the seat angle:

1. Remove the hardware that fastens the caster wheel to the caster fork.
2. Reposition the caster wheel on the caster fork.
3. Reinstall the hardware that fastens the caster wheel to the caster fork.
4. Repeat the procedure for the other side, making sure both caster wheels are adjusted to the same position.



**WARNING!** Changing the seat angle may adversely affect stability and cause the wheelchair to tip over.




Figure 38. Dual-post Armrest Height Adjustment




Figure 39. Caster Wheel Position

## IV. COMFORT ADJUSTMENTS

### Caster Angle Adjustment

When you change the seat angle, you must also change the caster angle so that the caster barrels remain at a 90° angle to the ground. See figure 40. There are 7 possible angles with a quick-release caster and 13 possible angles with a bolt-on caster.

**NOTE:** There are two screws that fasten the caster barrel assembly to the frame. The nuts that secure the screws are eccentric. The eccentric nuts should be positioned to the appropriate setting for the desired angle. A right-angle square is recommended for aligning the caster barrels to ensure 90° alignment to the ground.



**WARNING!** If your wheelchair is equipped with a transit securement system, the caster angle should only be adjusted by a Quantum Rehab Provider or a qualified technician.

#### To change the caster angle:

1. Remove the screws that fasten the caster wheel assembly to the frame.
2. Reposition the eccentric nuts to the desired setting and reinsert the bolts.
3. Tighten the screws.

**NOTE:** Make sure that both casters are adjusted to the same angle.

### Caster Wheel Removal

The casters can be removed one of two ways, depending on caster style.

#### To remove the bolt-on caster wheel:

1. Remove the hardware that attaches the caster wheel to the frame of your wheelchair. See figure 41.
2. Repeat for the other side if necessary.

#### To remove the quick-release caster wheel:

1. Press the quick-release button and pull the caster wheel off. See figure 42.
2. Repeat for the other side if necessary.




Figure 40. Caster Angle Adjustment




Figure 41. Bolt-on Caster Removal




Figure 42. Quick-release Caster Removal

## IV. COMFORT ADJUSTMENTS

### Positioning Belt

Your wheelchair may be equipped with a positioning belt that can be adjusted for operator comfort. See figure 43. The positioning belt is designed to support the operator so that he/she does not slide down or forward in the seat. The positioning belt is not designed for use as a restraining device in a motor vehicle.



**WARNING!** If your wheelchair is equipped with a transit securement system, the positioning belt should only be adjusted by a Quantum Rehab Provider or a qualified technician.

#### To install the positioning belt:

1. Remove and discard the screws on both the left and right back canes. See figure 43.

*NOTE: If your wheelchair is equipped with dual-post armrests, the existing back cane screws must be replaced with screws that are 2 in. (5 cm) long to properly install the positioning belt.*

2. Insert the longer screws through the positioning belt and then through the back canes on each side of the wheelchair seat. See figure 43.
3. Tighten both screws.
4. Install the positioning belt onto the bracket and pull until the positioning belt snaps into place.

#### To adjust the positioning belt for operator comfort:

1. Once seated, insert the metal tab on one side of the belt into the plastic housing on the opposite side until you hear a click. See figure 44.
2. Pull the excess strap attached to the metal tab until it is secure, but not so tight as to cause discomfort.

#### To release the positioning belt:

1. Push the button mechanism on the plastic housing.

**MANDATORY!** Make sure the positioning belt is properly secured to the wheelchair and is adjusted for operator comfort before each use.



**MANDATORY!** Inspect the positioning belt for loose or damaged parts before each use of the wheelchair. If you discover a problem, contact your Quantum Rehab Provider for maintenance and repair.




Figure 43. Positioning Belt Installation




Figure 44. Positioning Belt Adjustment

# V. CARE AND MAINTENANCE

## CARE AND MAINTENANCE

Your Litestream XF requires routine maintenance checks. You can perform some of these checks, but others require assistance from your Quantum Rehab Provider. Preventive maintenance is very important. If you follow the maintenance checks in this section as scheduled, you can help ensure that your wheelchair gives you years of trouble-free operation. If you have any doubt as to your wheelchair's care or operation, contact your Quantum Rehab Provider.



**WARNING! Do not service the wheelchair when the seat is occupied.**

Your wheelchair is susceptible to damage from the elements. Avoid damp areas of any kind.



**WARNING! Direct or prolonged exposure to water or dampness could cause the wheelchair to malfunction mechanically. Water can cause the chair's frame to rust. Wheelchairs should be examined periodically for signs of corrosion caused by water exposure, bodily fluids exposure, or incontinence. Damaged components should be replaced or treated immediately.**

### Should your wheelchair come in contact with water:

1. Dry your wheelchair as thoroughly as possible with a towel.
2. Allow your wheelchair to sit in a warm, dry place for 12 hours to allow unseen water to evaporate.
3. If any inconsistencies are found, take your wheelchair to your Quantum Rehab Provider. Wheelchairs that are frequently exposed to sources of water, such as incontinence, should be inspected often for corrosion.

### Temperature

Some of the parts of your wheelchair are susceptible to extreme changes in temperature. Always keep your wheelchair between the temperatures of 18° F (-8° C) and 122° F (50° C).

### General Guidelines

- Avoid prolonged exposure of your wheelchair to extreme conditions, such as heat, cold, or moisture.
- Make sure the pneumatic tires are inflated to the psi/bar/kPa air pressure rating indicated on the tire if equipped with pneumatic tires.



**WARNING! Make sure your tires are inflated to the psi/bar/kPa air pressure rating indicated on the tire. Do not underinflate or overinflate your tires. Low pressure may result in loss of control, and overinflated tires may burst. Overinflating tires can cause them to explode.**

**WARNING! Do not use a high pressure hose to inflate your tires.**

- Use a rubber conditioner on the tire sidewalls to help preserve them.



**WARNING! Never use a rubber conditioner on the tread area of the tires; doing so may make the tires slippery and cause your wheelchair to skid.**

- All wheel bearings are prelubricated and sealed. They require no subsequent lubrication.

### Daily Checks

- Check for loose or missing screws, nuts, and bolts. Replace or tighten them as necessary.
- Check the tires. Look for uneven wear, holes, or tire/rim separation. If they look excessively worn, then have them replaced.
- Visually inspect pneumatic tires for punctures.
- Confirm correct air pressure as marked on the side of the tire.
- Check for flat spots on solid tires. Flat spots could adversely affect stability.
- Confirm that the strap used to fold down the seatback is not torn or frayed.

# V. CARE AND MAINTENANCE

- Check the handrims. Make sure that they are fastened securely to the wheels.
- Check the wheel locks. Make sure they work properly. If they don't, instructions on how to adjust them are provided in this manual.



**WARNING! The wheel locks are designed to keep your chair stationary while you are parked. They are not designed to stop the wheelchair during driving. Do not use either of the wheel locks to stop the wheelchair while driving it. This can cause your wheelchair to become unstable and tip.**

## Weekly Checks

- Check for proper tire inflation. Pneumatic tires should be inflated to the psi/bar/kPa air pressure rating indicated on the tire. If a tire does not hold air, see your Quantum Rehab Provider for replacement of the tube.
- Check the anti-tip wheels for wear or damage. See your Quantum Rehab Provider for replacement or repair.

## Monthly Checks

- Check for drive tire wear. See your Quantum Rehab Provider for repair.
- Check the caster wheels for wear. Replace them as necessary.
- Check the caster forks for damage or fluttering which indicates that they may need to be adjusted or have the bearings replaced. See your Quantum Rehab Provider for repair.
- Keep your wheelchair clean and free of foreign material, such as mud, dirt, hair, food, drink, etc.

## Yearly Checks

Take your wheelchair to your Quantum Rehab Provider for yearly maintenance, especially if you use your wheelchair on a daily basis. This helps ensure that your wheelchair is functioning properly and helps prevent future complications.

## Storage

Your wheelchair should be stored in a dry place, free from temperature extremes.



**WARNING! If you fail to store the unit properly, the frame can rust.**

You may wish to place several boards under the frame of your wheelchair to raise it off of the ground during periods of prolonged storage. This takes the weight off the tires and reduces the possibility of flat spots developing on the areas of the tires contacting the ground.

## Disposal of Your Wheelchair

Your wheelchair must be disposed of according to applicable local and national statutory regulations. Contact your local waste disposal agency or Quantum Rehab Provider for information on proper disposal of wheelchair packaging, metal frame components, plastic components, neoprene, silicone, and polyurethane materials.

## Cleaning and Disinfection

- Use a damp cloth and mild, non-abrasive cleanser to clean the plastic and metal parts of your wheelchair. Avoid using products that may scratch the surface of your wheelchair.
- If necessary, clean your product with an approved disinfectant. Make sure the disinfectant is safe for use on your product before application.



**WARNING! Follow all safety instructions for the proper use of the disinfectant and/or cleaning agent before applying it to your product. Failure to comply may result in skin irritation or premature deterioration of upholstery and/or wheelchair finishes.**

**WARNING! Never hose off your wheelchair or place it in direct contact with water. Your wheelchair has a painted finish that allows it to be easily wiped clean with a damp cloth.**

# V. CARE AND MAINTENANCE



**WARNING!** Never use any chemicals to clean a vinyl seat, as they may cause the seat to become slippery or dry out and crack. Use soapy water and dry the seat thoroughly.

## Wheel Replacement

Your wheelchair may be equipped with a tire and tube, tire and flat-free insert, or a solid polyurethane tire. Replacement tires, tubes, inserts, and tire assemblies are readily available through your Quantum Rehab Provider.



**WARNING!** The wheels on your wheelchair should only be serviced or replaced by a Quantum Rehab Provider or a qualified technician.

**WARNING!** When changing a tire, remove the entire drive wheel. If any further disassembly is required, deflate the tire completely or it may explode.

**To replace a pneumatic tire tube, flat-free insert, or solid polyurethane tire:**

1. Remove the drive wheel from the frame. See figure 15.
2. If equipped with pneumatic tires, make sure that the tube is deflated completely.
3. Use several tire levers to pry the tire from the rim. Tire levers are available from a bicycle shop.

**NOTE:** If your wheelchair is equipped with solid polyurethane tires, then no further disassembly is required. Go to step 7.

4. Remove the old tube or flat-free insert.
5. If equipped with pneumatic tires, inflate the new tube with enough air so that it just starts to take shape.
6. If equipped with pneumatic tires or flat-free inserts, install the tube or flat-free insert into the tire.
7. Use the tire levers to secure the tire back onto the rim.
8. If equipped with pneumatic tires, inflate the tire to the psi/bar/kPa air pressure recommended on the sidewall. Always use a regulated air source.
9. Reinstall the drive wheel to the frame. Make sure that the wheel axle is snapped securely into the frame bracket.

## Manual Wheel Locks

After several months of use, you may find it necessary to adjust the wheel locks so that they lock securely when engaged. You also may need to adjust them after changing the drive wheel position. Your wheelchair may be equipped with push-to-lock, pull-to-lock, or scissor-style wheel locks. Adjustment instructions for each are outlined below.



**WARNING!** If your wheelchair is equipped with pneumatic tires, make sure they are inflated to the pressure recommended on the sidewall of the tire.

**To adjust each push-to-lock or pull-to-lock manual wheel lock:**

1. Disengage the wheel lock. See figure 6. The wheel lock should be 0.25 in. (0.64 cm) away from the tire for both pneumatic and solid tires.
2. Loosen the hardware that fastens the wheel lock to the frame. See figure 45.
3. Reposition the wheel lock so that it is 0.25 in. (0.64 cm) away from the tire and perpendicular to the plane of the tire.
4. Tighten the hardware that fastens the wheel lock to the frame.
5. Engage the wheel lock and note the position of the blade.
6. Reposition the wheel lock if necessary.
7. Repeat the procedure for the other side.




Figure 45. Wheel Lock Adjustment

# V. CARE AND MAINTENANCE



**WARNING! A Quantum Rehab Provider should adjust the wheel lock with the occupant in the wheelchair on the maximum incline to ensure that the wheel lock catches properly.**

## To adjust each scissor-style wheel lock:

1. Disengage the wheel lock. The front of the wheel lock extrusion that faces the caster wheels should be exactly 2.1 in. (5.3 cm) from where the wheel lock blade makes contact with the plane of the drive wheel. See figure 46. This applies to both solid and pneumatic tires.
2. Loosen the hardware that fastens the wheel lock to the frame. See figure 46.
3. Reposition the wheel lock so that the front of the wheel lock extrusion that faces the caster wheels is exactly 2.1 in. (5.3 cm) from where the wheel lock blade makes contact with the plane of the drive wheel. This applies to both solid and pneumatic tires.
4. Tighten the hardware that fastens the wheel lock to the frame.
5. Repeat the procedure for the other side.

## Seat Tension

After a period of use, you may find it necessary to adjust the seat tension.

## To adjust the seat tension:

1. Loosen the screws that attach the seat to the side rail on the left side (side with hook and loop fastener). See figure 47.
2. Unfasten the hook and loop fastener.
3. Grab the bottom flap and pull the seat tight.
4. Fasten the hook and loop fastener.
5. Tighten the screws.

## When to See Your Quantum Rehab Provider for Service

The following symptoms could indicate a serious problem with your wheelchair. If necessary, contact your Quantum Rehab Provider. When calling, have the model number, serial number, and nature of the problem.

- Uneven wear on any of the tires
- Bent or broken wheel assemblies
- Loose or missing hardware that attaches the transit securement brackets or positioning belt anchors to the wheelchair.




Figure 46. Scissor-style Wheel Lock Adjustment




Figure 47. Seat Tension Adjustment

## **VI. WARRANTY**

Pride warrants this product, so far as the same is of its own manufacture, to be free of defects in material and workmanship, when used by the purchaser under normal conditions, for (1) year after purchase. The side frames and cross-frames are warranted for the life of the product to the original purchaser.

The warranties herein specified shall not apply if the failure of any component of this product to satisfy such warranty is due to accident, neglect, misuse, improper transportation arranged by the purchaser or its agent, or causes other than ordinary use including operation contrary to Pride's operating instructions or if any person other than Pride's authorized personnel shall modify, adjust, or repair such product or perform any related service on it.

It must be understood that the purchaser assumes full responsibility for the overall safe operation in the environment for which the product is intended to function.

# APPENDIX I - SPECIFICATIONS

<b>LITESTREAM XF SPECIFICATIONS (WITHOUT DEPTH-ADJUSTABLE SIDE FRAME)</b>	
Seat Dimensions:	Width: 14-21 in.; 18-22 in. with HD option Depth: 14-20 in.; 18-20 in. with HD option
Back Height:	15-19 in.; 12-16 in.; 15-19 in. with HD option
Seat and Seatback:	Adjustable tension
Drive Wheels:	Mags: 20 in. (50.80 cm), 22 in. (55.88 cm), and 24 in. (60.96 cm) Spoke: 20 in. (50.80 cm), 22 in. (55.88 cm), 24 in. (60.96 cm), 25 in. (63.5 cm), and 26 in. (66.04 cm) Tuff Wheel®: 24 in. (60.96 cm) (HD option only) Sport Light Extreme: 24 in. (60.96 cm), 25 in. (63.5 cm), and 26 in. (66.04 cm) Light Extreme: 22 in. (55.88 cm), 24 in. (60.96 cm), 25 in. (63.5 cm), and 26 in. (66.04 cm)
Caster Wheels:	Solid: 4 in. (10.16 cm), 5 in. (12.70 cm), 6 in. (15.24 cm), or 8 in. (20.32 cm) Pneumatic: 6 in. (15.24 cm) <sup>3</sup> and 8 in. (20.32 cm)
Axle Mount:	Quick-release
Caster Fork:	Angle-adjustable, Quick-release, Bolt-on, Frog Legs™
Maximum Safe Slope:	5° (8.7%)
Turning Radius: <sup>1</sup>	<b>See figure 48.</b>
Overall Size: <sup>1</sup>	Length: <b>See figure 48.</b> Width: <b>See figure 48.</b>
Weight Capacity:	250 lbs./113.4 kg/18 stone 350 lbs./158.7 kg/25 stone with HD option
Weights:	27.3 lbs. (12.38 kg) without front riggings 35.5 lbs. (16.1 kg) without front riggings (HD option)
Armrests:	Desk-length (tubular); Height-adjustable T-arm, Dual-post
Front Rigging:	Swing-away leg rests, Elevating leg rests, Adjustable foot plates
Wheel Locks:	Push-to-lock; Pull-to-lock; Scissor-lock
Accessories:	Anti-tip wheels
Static Stability Forward (Wheels Locked):	7.5°
Static Stability Rearward (Wheels Locked):	8.5°
Wheel Lock Operational Force: <sup>2</sup>	Left: 60 N Right: 60 N

<sup>1</sup> Due to manufacturing tolerances and continued product improvement, this specification can be subject to a variance of (+ or -) 3%.

<sup>2</sup> Test results for Litestream XF equipped with push-to-lock wheel locks in accordance with ANSI-RESNA WC Vol. 2 Section 03. Maximum operating force for hand operated wheel locks is 60N (+ or -) 10%. Operational force can vary based on tire type, manufacture, and condition.

<sup>3</sup> Pneumatic caster wheels are intended for indoor/outdoor use on smooth, level surfaces. They are not recommended for use on rugged terrain as this may cause premature wear and more frequent replacement of the tire.

**NOTE: This product conforms to all applicable ANSI-RESNA testing requirements and ISO 7176 series EN12183 standards. All specifications subject to change without notice.**

# APPENDIX I - SPECIFICATIONS

LITESTREAM XF SPECIFICATIONS (WITH DEPTH-ADJUSTABLE SIDE FRAME)	
Seat Dimensions:	Width: 14-21 in. Depth: 14-20 in.; 3 in. (7.62 cm) of depth adjustment
Back Height:	16 - 24 in.
Seat and Seatback:	Seat: Adjustable tension; Seatback: Synergy® Shape Back or similar back
Drive Wheels:	Mags: 20 in. (50.80 cm), 22 in. (55.88 cm), and 24 in. (60.96 cm) Spoke: 20 in. (50.80 cm), 22 in. (55.88 cm), 24 in. (60.96 cm), 25 in. (63.5 cm), and 26 in. (66.04 cm)
Caster Wheels:	Solid: 4 in. (10.16 cm), 5 in. (12.70 cm), 6 in. (15.24 cm), or 8 in. (20.32 cm) Pneumatic: 6 in. (15.24 cm) <sup>3</sup> and 8 in. (20.32 cm)
Axle Mount:	Quick-release
Caster Fork:	Angle-adjustable, Quick-release, Bolt-on, Frog Legs™
Maximum Safe Slope:	5° (8.7%)
Turning Radius: <sup>1</sup>	<b>See figure 48.</b>
Overall Size: <sup>1</sup>	Length: <b>See figure 48.</b> Width: <b>See figure 48.</b>
Weight Capacity:	250 lbs./113.4 kg/18 stone
Weights:	29.5 lbs. (13.38 kg) without front riggings
Armrests:	Latching flip-up; Non-latching flip-up; Height-adjustable T-arm
Front Rigging:	Swing-away leg rests, Elevating leg rests, Adjustable foot plates
Wheel Locks:	Push-to-lock; Pull-to-lock; Scissor-lock
Accessories:	Anti-tip wheels
Static Stability Forward (Wheels Locked):	7.5°
Static Stability Rearward (Wheels Locked):	8.5°
Wheel Lock Operational Force: <sup>2</sup>	Left: 60 N Right: 60 N

<sup>1</sup> Due to manufacturing tolerances and continued product improvement, this specification can be subject to a variance of (+ or -) 3%.

<sup>2</sup> Test results for Litestream XF equipped with push-to-lock wheel locks in accordance with ANSI-RESNA WC Vol. 2 Section 03. Maximum operating force for hand operated wheel locks is 60N (+ or -) 10%. Operational force can vary based on tire type, manufacture, and condition.

<sup>3</sup> Pneumatic caster wheels are intended for indoor/outdoor use on smooth, level surfaces. They are not recommended for use on rugged terrain as this may cause premature wear and more frequent replacement of the tire..

**NOTE: This product conforms to all applicable ANSI-RESNA testing requirements and ISO 7176 series EN12183 standards. All specifications subject to change without notice.**

# APPENDIX I - SPECIFICATIONS








Figure 48. Litestream XF Dimensions and Ground Clearance





# APPENDIX II - STF HEIGHT MATRIX

LITESTREAM XF SEAT-TO-FLOOR/AXLE PLATE MOUNTING MATRIX				
Axle Plate Mounting Location	20 in. Wheel	22 in. Wheel	24 in. Wheel **	26 in. Wheel
	12.5 in. (31.75 cm)	13.5 in. (34.29 cm)	14.5 in. (36.83 cm)	15.5 in. (39.37 cm)
	13 in. (33.02 cm)	14 in. (35.56 cm)	15 in. (38.1 cm)	16 in. (40.64 cm)
	13.5 in. (34.29 cm)	14.5 in. (36.83 cm)	15.5 in. (39.37 cm)	16.5 in. (41.91 cm)
	14 in. (35.56 cm)	15 in. (38.1 cm)	16 in. (40.64 cm)	17 in. (43.8 cm)

\*\* High performance 100 psi tires result in 0.5 in. (1.27 cm) lower STF.

**NOTE:** Extreme posterior seat angle (over 2 in. [5 cm] deviation from front to rear STF) will result in a STF height of up to 0.25 in. (0.64 cm) less than what is stated.

# APPENDIX II - STF HEIGHT MATRIX

LITESTREAM XF SEAT-TO-FLOOR/AXLE PLATE MOUNTING MATRIX				
Axle Plate Mounting Location	20 in. Wheel	22 in. Wheel	24 in. Wheel **	26 in. Wheel
	14.5 in. (36.83 cm)	15.5 in. (39.37 cm)	16.5 in. (41.91 cm)	17.5 in. (44.45 cm)
	15 in. (38.1 cm)	16 in. (40.64 cm)	17 in. (43.8 cm)	18 in. (45.72 cm)
	15.5 in. (39.37 cm)	16.5 in. (41.91 cm)	17.5 in. (44.45 cm)	18.5 in. (46.99 cm)
	16 in. (40.64 cm)	17 in. (43.8 cm)	18 in. (45.72 cm)	19 in. (48.26 cm)

\*\* High performance 100 psi tires result in 0.5 in. (1.27 cm) lower STF.

**NOTE:** Extreme posterior seat angle (over 2 in. [5 cm] deviation from front to rear STF) will result in a STF height of up to 0.25 in. (0.64 cm) less than what is stated.

# APPENDIX II - STF HEIGHT MATRIX

LITESTREAM XF CASTER WHEEL FRONT SEAT-TO-FLOOR HEIGHT MATRIX				
STF Height	Caster Size			
	4 in.	5 in.	6 in.	8 in.
12.5 in.	2US	-	-	-
13 in.	1US, 3UF	2US	-	-
13.5 in.	3UL, 2UF	1US, 4UL	-	-
14 in.	2UL, 1UF	3UL, 2UF	1US, 4UL	-
14.5 in.	1UL	2UL, 1UF	3UL	-
15 in.	3S	1UL	2UL, 1UF	-
15.5 in.	2S	-	1UL	-
16 in.	1S, 3F	2S	-	2UL
16.5 in.	3L, 2F	1S, 4L	-	1UL
17 in.	2L, 1F	3L, 2F	1S, 4L	-
17.5 in.	1L	2L, 1F	3L	-
18 in.	-	1L	2L, 1F	-
18.5 in.	-	-	1L	-
19 in.	-	-	-	2L
19.5 in.	-	-	-	1L

S = Short Fork (FRMASMB9265)

US= Ultra Hemi with Short Fork

L = Long Fork (FRMASMB9264)

UL = Ultra Hemi with Long Fork

F = Frog Legs™

UF - Ultra Hemi with Frog Legs™

1, 2, 3, 4 = Hole position from bottom




Figure 49. Long and Short Caster Forks

# NOTES

## ***Quality Control - Litestream XF***

**Inclusion of all Parts**

**Fit and Finish**

**Performance**



**Pride keeps a more detailed report on file at the factory.**

\_\_\_\_\_  
**Date Inspected**

\_\_\_\_\_  
**Inspector**

**\* I NFMANU2788 \***

\_\_\_\_\_  
**Product Serial #**